

<b>AMENDMENT OF SOLICITATION/MODIFICATION OF CONTRACT</b>			1. CONTRACT ID CODE	PAGE OF PAGES 1   2
2. AMENDMENT/MODIFICATION NO. A00001	3. EFFECTIVE DATE Jan 24, 2013	4. REQUISITION/PURCHASE REQ. NO. 53-4014-13	5. PROJECT NO. (If applicable)	
6. ISSUED BY Supply Officer (Code 3410) Naval Research Laboratory 4555 Overlook Ave SW Washington DC 20375-5329	CODE	7. ADMINISTERED BY (If other than Item 6) Purchasing Officer (Code 3410) Naval Research Laboratory 4555 Overlook Ave SW Washington DC 20375-5329	CODE	N00173
8. NAME AND ADDRESS OF CONTRACTOR (No., street, county, State and ZIP Code) All Quoters			(X)	9A. AMENDMENT OF SOLICITATION NO. N00173-13-Q-0080
				9B. DATED (SEE ITEM 11) Jan 17, 2013
				10A. MODIFICATION OF CONTRACT/ORDER NO.
				10B. DATED (SEE ITEM 13)
CODE	FACILITY CODE			

**11. THIS ITEM ONLY APPLIES TO AMENDMENTS OF SOLICITATIONS**

The above numbered solicitation is amended as set forth in Item 14. The hour and date specified for receipt of Offers  is extended,  is not extended.

Offers must acknowledge receipt of this amendment prior to the hour and date specified in the solicitation or as amended, by one of the following methods:

- (a) By completing items 8 and 15, and returning 1 copies of the amendment; (b) By acknowledging receipt of this amendment on each copy of the offer submitted; or (c) By separate letter or telegram which includes a reference to the solicitation and amendment numbers. FAILURE OF YOUR ACKNOWLEDGMENT TO BE RECEIVED AT THE PLACE DESIGNATED FOR THE RECEIPT OF OFFERS PRIOR TO THE HOUR AND DATE SPECIFIED MAY RESULT IN REJECTION OF YOUR OFFER. If by virtue of this amendment your desire to change an offer already submitted, such change may be made by telegram or letter, provided each telegram or letter makes reference to the solicitation and this amendment, and is received prior to the opening hour and date specified.

12. ACCOUNTING AND APPROPRIATION DATA (If required)

**13. THIS ITEM ONLY APPLIES TO MODIFICATION OF CONTRACTS/ORDERS. IT MODIFIES THE CONTRACT/ORDER NO. AS DESCRIBED IN ITEM 14.**

CHECK ONE	A. THIS CHANGE ORDER IS ISSUED PURSUANT TO: (Specify authority) NO. IN ITEM 10A.	THE CHANGES SET FORTH IN ITEM 14 ARE MADE IN THE CONTRACT ORDER
<input type="checkbox"/>	B. THE ABOVE NUMBERED CONTRACT/ORDER IS MODIFIED TO REFLECT THE ADMINISTRATIVE CHANGES (such as changes in paying office, appropriation date, etc.) SET FORTH IN ITEM 14, PURSUANT TO THE AUTHORITY OF FAR 43.103(b).	
<input type="checkbox"/>	C. THIS SUPPLEMENTAL AGREEMENT IS ENTERED INTO PURSUANT TO AUTHORITY OF:	
<input type="checkbox"/>	D. OTHER (Specify type of modification and authority)	

**E. IMPORTANT:** Contractor  is not,  is required to sign this document and return \_\_\_\_\_ copies to the issuing office.

14. DESCRIPTION OF AMENDMENT/MODIFICATION (Organized by UCF section headings, including solicitation/contract subject matter where feasible.)

The above referenced Request for Quotation (RFQ) is amended to answer a Contractors questions and to extend the closing date to 1-31-13.

See Attachment 1 for questions and answers.

All other terms and conditions remain the same.

Except as provided herein, all terms and conditions of the document referenced in Item 9A or 10A, as heretofore changed, remains unchanged and in full force and effect.

15A. NAME AND TITLE OF SIGNER (Type or print)	16A. NAME AND TITLE OF CONTRACTING OFFICER (Type or print)
	Ruth Dixon
15B. CONTRACTOR/OFFEROR	15C. DATE SIGNED
(Signature of person authorized to sign)	
16B. UNITED STATES OF AMERICA	16C. DATE SIGNED
(Signature of Contracting Officer)	1/24/13

- 1) What is the frequency range are you measurements? DC to 100GHz
- 2) What size are your wafers? 8 inch max
- 3) Are you planning to probe wafers, pieces, packages, etc.? YES, all.
- 4) What is the lowest current and voltage you need to measure? This procurement activity does not address precision dc measurement equipment.
- 5) What is the highest current and voltage you need to measure? DC probes should be capable of handling 200V and up to 15Amps
- 6) What is the smallest feature size you need to visually resolve? 100 nm
- 7) What is the smallest feature size you need to touch with the probes? 10um